

Armstrong Medical Ltd

Wattstown Business Park, Newbridge Road Coleraine, BT521BS Northern Ireland

10/05/2023

Confirmation Letter Reference: CLNB1639 GBPC08349

To whom it may concern,

Confirmation of receipt of a formal application and conclusion of written agreement in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices

This letter confirms that, SGS Belgium NV, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1639 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Armstrong Medical Ltd

Wattstown Business Park, Newbridge Road Coleraine, BT521BS
Northern Ireland
SRN Number: XI-IM-000004090

The devices covered by the formal application and the written agreement mentioned above are shown at the end of this letter.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired prior to the publication of the Regulation EU 2023/607 on 15 March 2023, this letter also confirms that:

- the manufacturer submitted the MDR application and signed the written agreement by the date of MDD/AIMDD certificate expiry;
- the certificates expired after 26 May 2021 by course of time and were valid at the date of their expiry neither having been suspended nor withdrawn.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3 of MDR (as amended by EU 2023/607), are shown below:

26 May 2026 for Class III custom-made implantable devices

SGS Belgium NV

Certification and Business Enhancement Registered Office: Noorderlaan 87 BE-2030 Antwerpen t +32 (0)3 545 48 48 f +32 (0)3 545 48 49 Boulevard International/Internationalelaan 55D BE-1070 Brussels t+32 (0)2 556 00 40 f +32 (0)3 545 48 49 www.be.sgs.com



- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Wellestablished technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body SGS Belgium NV 1639,

pp [Sean Kelly]

Virginie SILORET

Global Medical Device Certification Manager

Email: <u>Virginie.siloret@sgs.com</u> Phone: +41 22 739 98 58

Devices covered by this letter:

Device name / Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre- application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Anaesthetic Circuits;	Class IIa	N/A	GB19/964541;
Ventilator Circuits; CPAP			NB1639
Circuits;			
506014624TF012S			
.63			
Resuscitation Sets;			
506014624TF042Y			
Humidification Chambers,			
506014624TF1535			
Drug Nebulizers;			
506014624TF0634			
Catheter Mounts;			
506014624TF122X			
Respiratory Face Masks;			

SGS Belgium NV

Certification and Business Enhancement Registered Office: Noorderlaan 87 BE-2030 Antwerpen t +32 (0)3 545 48 48 f +32 (0)3 545 48 49 Boulevard International/Internationalelaan 55D BE-1070 Brussels t+32 (0)2 556 00 40 f +32 (0)3 545 48 49 www.be.sgs.com



Device name / Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre- application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
506014624TF2232			A
Water Traps;			(1)
506014624TF0838			
Breathing Bags;			0,2
506014624TF0838			
Gas Sampling Lines;			120
506014624TF0838			
Carbon Dioxide Absorbents;			79 ,
506014624TF1433			
506014624TF1434			
506014624TF1435			
506014624TF1436		10	
506014624TF1437			
Absorbent Canister Adapters,		000	
506014624TF193D			
Nasal Interfaces,		¥ .	
506014624TF2334	Citte		
Gas Flow Driver	Class IIb	N/A	GB19/964541;
506014624TF2436	Hion,		NB1639

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action
2023/05/10	Version 1	Initial issue